AS ONE offers everything you need from laboratory instruments and consumables, factory MRO, to hospital & nursing care supplies

AS ONE Corporation

1-5486-21 Auto Dry Desiccator 330 x 345 x 525mm OH-3S



Order No.	1-5486-21
Model No.	OH-3S
JAN Code	4560111767828
	JPY: 46,000
Standard price	USD: 321.95
	Excange rate 1USD= 142.88JPY Valid price in Japan
Quantity	1unit









Features

- Dehumidification unit, which does not incorporate driving device, does not generate sound or vibration, and is very durable.
- Energy-saving design has been adopted, consume only 3W, and work as economical dehumidification storage cabinets that are always turned on.
- Moisture inside cabinet is released directly in gaseous state.
- <Light-shielding Type>
- Suitable for storage of samples sensitive to light, as 5 sides other than door are made of lightshielding materials.

Spec

- Auto dry type
- Material: Transparent PMMA (acrylic)
- Dehumidification system: Using solid high polymer electrolyte membrane, H2O electrolytic discharge
- Reaching humidity: 25%RH (with no internal load, depending on the usage environment)
- Dehumidifying element membrane area: 30 x 30mm
- Power supply: AC100-240V 50/60Hz
- Power consumption: 3W
- Shelf board interval: 1 stage in the bottom/50mm, Other/30mm
- Weight: 5kg
- Power cord length: 2m (2P plug)
- Accessories: Shelf board x 3, Low humidity digital thermo-hygrometer (THI-HP), silica gel plate x 1
- Model number: OH-3S
- Outer dimension (mm) *: 330 x 345 x 525 (*including dehumidifying unit.)
- Inner dimension (mm): 285 x 275 x 485
- Country of origin: JAPAN
- Control Panel: N/A
- Plug: Type A(https://en.wikipedia.org/wiki/Mains_electricity_by_country)
- RoHS3 (EU)2015/863 Certificates can be issued
- ▶ Package size: 420×400×635 mm 6.8 kg

Catalog Name	Page
AS ONE Catalog 2023 [Instruments for Laboratory]	1170
AS ONE Catalog 2021 [Instruments for Laboratory]	1116
AS ONE Catalog 2019 [Instruments for Laboratory]	1097
AS ONE Catalog 2017 [Instruments for Laboratory]	863
AS ONE Catalog 2015 [Instruments for Laboratory]	803